

CERTIFICATE OF ANALYSIS

Article no. 1448
Batch no. P2I718032L

Methanol for residue analysis (min. 99.9 %)
CH₃OH

For laboratory use.

Parameter	Specification	Result
Assay		100.00 %
Refractive index (20 °C)	1.327 - 1.331	passes test
Water (KF)	max. 500 ppm	passes test
Non-volatile substances	max. 5 ppm	passes test
Free acid (as HCOOH)	max. 10 ppm	passes test
Free alkali (as NH ₃)	max. 1 ppm	passes test
GC-ECD: Peak (lindane) (retention time trichlorobenzene to mirex)	max. 3 ng/l	passes test
GC-NPD: Peak (ethylparathion) (retention time atrazin to coumaphos)	max. 3 ng/l	passes test
Colour (Hazen)	max. 10	passes test

Test date: 06.10.2022
Use before date: 10-2024

This document has been generated automatically and is valid without signature.