



Agilent 6530B Accurate Mass Q-TOF LC/MS System

Data Sheet



Parameter	Measure	Specification
Sensitivity, MS mode, electrospray on-column, 400 $\mu\text{L}/\text{min}$ flow rate	1 pg LC/MS injection of reserpine signal-to-noise for the reserpine (M+H) ⁺ at m/z 609.2807 while maintaining a resolution of 20,000 at m/z 1,522 in 4 GHz mode	200:1 RMS Option 200: 60:1 RMS
Sensitivity, MS/MS mode, electrospray on-column, 400 $\mu\text{L}/\text{min}$ flow rate	1 pg LC/MS injection of reserpine signal-to-noise for most intense product ions (174, 195, 397, 448 m/z) while maintaining a resolution of 20,000 at m/z 1,522 in 4 GHz mode	500:1 RMS Option 200: 180:1 RMS
Mass resolving power	Measured at m/z 322 after automatic tuning procedure Measured at m/z 1,522 after automatic tuning procedure	Greater than 10,000 at 118 m/z Greater than 20,000 FWHM at 1,522 m/z
Mass accuracy, MS mode, electrospray on-column, 400 $\mu\text{L}/\text{min}$	Measured at the (M+H) ⁺ ion of reserpine (m/z 609.2807) using an internal mass reference	Better than 2 ppm RMS as measured from 10 repeat injections
Mass accuracy, MS/MS mode, electrospray on-column, 400 $\mu\text{L}/\text{min}$	Product ion 397 m/z for reserpine	Better than 3 ppm RMS on m/z 397 as measured from 10 repeat injections
Mass accuracy temperature stability, MS mode	Temperature: 15 to 35 °C (59 to 95 °F) at constant temperature	Maintain 1 ppm mass accuracy (variations < 3 °C from calibration temperature)



Agilent Technologies

Parameter	Measure	Specification
Dynamic range	Intrascan dynamic range on co-eluting components	Up to 5 decades
Mass range		m/z 100 – 20,000 extended mass range m/z 50 – 1,700 or 50 – 3,200 for both high resolution and extended dynamic range modes Quadrupole up to 4,000 m/z
Spectral acquisition rate, MS mode	m/z 50 to 1,700 in MS mode while maintaining a resolution of 20,000 at m/z 1,522 in 4 GHz mode	50 spectra/second
Spectral acquisition rate, MS/MS mode	m/z 50 to 1,700 in MS/MS mode while maintaining a resolution of 20,000 at m/z 1,522 in 4 GHz mode	50 MS/MS spectra/second
Positive to negative switching	Complete cycle switching from positive to negative and positive modes – allows for stabilization time	1.5 seconds

All specifications are achieved in manufacturing, and instrument performance data is supplied with shipment. All specification values are achieved after autotune and do not require manual optimization. These specifications are not standard installation specifications for the Agilent 6530B Accurate Mass Q-TOF LC/MS System. Agilent high resolution accurate mass Q-TOF instruments are tested and installed in accordance with standard performance tests as described in the Agilent installation manual.

www.agilent.com/chem/QTOF

This information is subject to change without notice.

© Agilent Technologies, Inc., 2012
Published in the USA, November 9, 2012
5991-1422EN



Agilent Technologies